



## Ohio Administrative Code

### Rule 3701:1-58-69 Five-year inspection for teletherapy and gamma stereotactic radiosurgery units.

Effective: August 15, 2021

---

(A) A licensee shall have each teletherapy unit and gamma stereotactic radiosurgery unit fully inspected and serviced during source replacement, to assure proper functioning of the source exposure mechanism and other safety components. The interval between each full-inspection servicing shall not exceed five years for each teletherapy unit and shall not exceed seven years for each gamma stereotactic radiosurgery unit.

(B) This inspection and servicing may only be performed by persons specifically licensed to do so by the director, the United States nuclear regulatory commission, or an agreement state.

(C) A licensee shall keep a record of the inspection and servicing in accordance with rule 3701:1-58-100 of the Administrative Code.

---